



THE SOUTH INDIA TEXTILE RESEARCH ASSOCIATION

CENTRE OF EXCELLENCE FOR MEDICAL TEXTILES
Physical, Chemical & Biological Testing Laboratories

(ISO/IEC 17025 NABL ACCREDITED)

13/37, Avinashi Road, Aerodrome Post, Coimbatore - 641 014, INDIA



Cert. Number:TC-6944

CoE - PHYSICAL LABORATORY

XXVII / P / E / 7544 / 20

ULR: TC694420600003943P

To

Frabon India Tech Solutions

107,Abhijeet 1,Mithakhali Six Roads Circle

Mithakhali

Ahmedabad,Gujarat ,India

Pincode : 380009

Contact No(s) : 9265298784

Email : cbo@frabon.com

Inward / Report No : P2005329

Inward Date : 22-05-2020

Report Date : 25-05-2020

Dear Sir / Madam,

This has reference to the sample(s) submitted by you for testing vide your letter reference no :
Letter dt 13.05.2020.

The results pertaining to your sample(s) are enclosed herewith.

Encl : Bill

This is a computer generated report, hence does not require signature.

IMPORTANT

This report is strictly CONFIDENTIAL. Its use for publicity, arbitration or as evidence legal disputes is forbidden. Reference of sample(s) given by the party. The above results are related to the samples tested. The report shall not be reproduced except in full, without the written approval of the laboratory.



THE SOUTH INDIA TEXTILE RESEARCH ASSOCIATION

CENTRE OF EXCELLENCE FOR MEDICAL TEXTILES
Physical, Chemical & Biological Testing Laboratories

(ISO/IEC 17025 NABL ACCREDITED)

13/37, Avinashi Road, Aerodrome Post, Coimbatore - 641 014, INDIA



Cert. Number:TC-6944

Frabon India Tech Solutions, Ref : Letter dt 13.05.2020

ULR: TC69442060003943P

(Sample Tested at : R.H.65% +/- 2% and Temp. 21 Degree C +/- 1 Degree C)

Differential Pressure** IS 16289:2014 Annexure C Pa/cm ²	P2005329-1 Described by the Customer : Y-Cure Face Mask 39.0
Splash Resistance ASTM F1862/F1862M-13 Result at 160mmHg	P2005329-1 Described by the Customer : Y-Cure Face Mask Pass
Particulate Filtration Efficiency at 0.3 Microns ASTM F2299/F2299M-03 (2010) Percentage	P2005329-1 Described by the Customer : Y-Cure Face Mask 97.45